



RLS BATCH CERTIFICATE

Certificate Number	AI/3743/BLF1000
Test Method	Gravimetric method based on DIN/ISO 3819
Item Tested	Beaker
Catalogue Number	2060420_10
Batch Number	09.20
Item Tested On	01/09/2020
Material Used	3.3 Borosilicate Glass
Scale Used For Calibration	Sartorius 31610xx, LC 0.01g
Thermometer Used for Calibration	Greisinger TS01, LC 0.1 °C

WARRANTY CERTIFICATE AT 20 °C

Certified Volume	1000.15mL
Norminal Volume	1000mL
Tolerance	+ -0.12

The above product has been examined for compliance to DIN/ISO, Class A. It has been found to satisfy the specified requirements. The above item has been calibrated at 20 °C at RH 65%. The above calibration result is valid at the time of calibration under lab conditions.