



RLS BATCH CERTIFICATE

| | |
|---|---|
| Certificate Number | AI/3743/BLF10 |
| Test Method | Gravimetric method based on DIN/ISO 3819 |
| Item Tested | Beaker |
| Catalogue Number | 2060420_1 |
| Batch Number | 09.20 |
| Item Tested On | 01/09/2020 |
| Material Used | 3.3 Borosilicate Glass |
| Scale Used For Calibration | Sartorius 31610xx, LC 0.01g |
| Thermometer Used for Calibration | Greisinger TS01, LC 0.1 °C |

WARRANTY CERTIFICATE AT 20 °C

| | |
|-------------------------|----------------|
| Certified Volume | 10.01mL |
| Norminal Volume | 10mL |
| Tolerance | + -0.02 |
| | |
| | |

The above product has been examined for compliance to DIN/ISO, Class A. It has been found to satisfy the specified requirements. The above item has been calibrated at 20 °C at RH 65%. The above calibration result is valid at the time of calibration under lab conditions.