



## RLS BATCH CERTIFICATE

<b>Certificate Number</b>	<b>AI/3743/BLF250</b>
<b>Test Method</b>	<b>Gravimetric method based on DIN/ISO 3819</b>
<b>Item Tested</b>	<b>Beaker</b>
<b>Catalogue Number</b>	<b>2060420_6</b>
<b>Batch Number</b>	<b>09.20</b>
<b>Item Tested On</b>	<b>01/09/2020</b>
<b>Material Used</b>	<b>3.3 Borosilicate Glass</b>
<b>Scale Used For Calibration</b>	<b>Sartorius 31610xx, LC 0.01g</b>
<b>Thermometer Used for Calibration</b>	<b>Greisinger TS01, LC 0.1 °C</b>

WARRANTY CERTIFICATE AT 20 °C

<b>Certified Volume</b>	<b>250.01mL</b>
<b>Norminal Volume</b>	<b>250mL</b>
<b>Tolerance</b>	<b>+ -0.12</b>

The above product has been examined for compliance to DIN/ISO, Class A. It has been found to satisfy the specified requirements. The above item has been calibrated at 20 °C at RH 65%. The above calibration result is valid at the time of calibration under lab conditions.