



## RLS BATCH CERTIFICATE

|   |   |
|---|---|
| <b>Certificate Number</b>               | <b>AI/3743/BLF600</b>                           |
| <b>Test Method</b>                      | <b>Gravimetric method based on DIN/ISO 3819</b> |
| <b>Item Tested</b>                      | <b>Beaker</b>                                   |
| <b>Catalogue Number</b>                 | <b>2060420_9</b>                                |
| <b>Batch Number</b>                     | <b>09.20</b>                                    |
| <b>Item Tested On</b>                   | <b>01/09/2020</b>                               |
| <b>Material Used</b>                    | <b>3.3 Borosilicate Glass</b>                   |
| <b>Scale Used For Calibration</b>       | <b>Sartorius 31610xx, LC 0.01g</b>              |
| <b>Thermometer Used for Calibration</b> | <b>Greisinger TS01, LC 0.1 °C</b>               |

WARRANTY CERTIFICATE AT 20 °C

|                         |                 |
|-------------------------|-----------------|
| <b>Certified Volume</b> | <b>600.05mL</b> |
| <b>Norminal Volume</b>  | <b>600mL</b>    |
| <b>Tolerance</b>        | <b>+ -0.12</b>  |
|                         |                 |
|                         |                 |
|                         |                 |

The above product has been examined for compliance to DIN/ISO, Class A. It has been found to satisfy the specified requirements. The above item has been calibrated at 20 °C at RH 65%. The above calibration result is valid at the time of calibration under lab conditions.