

CERTIFICATE OF ANALYSIS[®]

CHEMICAL NAME
CHEMICAL GRADE
CAS NUMBER
APPEARANCE
LOT NUMBER
EXPIRY DATE

POTASSIUM AA STANDARD 1000 ppm (in 0.5N HNO₃)
LABORATORY GRADE
7440-09-7
COLOURLESS LIQUID
RLS-18724-3
18/07/2026

Warranty Certificate

<i>Test</i>	<i>Specifications</i>	<i>Results</i>
ASSAY (%)		1000ppm
APPEARANCE	Colorless	Within the specification
DENSITY AT 25 °C (g/cm³)	1,0 -1.1	1.0
Concentration of HNO₃ (N)	0.1-0.5	0.5

RLS Chemical warrants conformance of its products against the relevant specification in good faith. In-house quality control procedures include representations provided by production/manufacturing source. A representative batch retention sample of the above-mentioned material has tested according to results above. Shelf life is based on current knowledge and holds only for correct storage in the original closed and sealed containers. Suitability of the product for a particular use or application is not guaranteed. It remains the responsibility of the user to determine such.

APPROVED DATE: 14.07.2024



Quality Control

Quality Management ISO 9001:2015
NIST traceable calibration certified.