

CERTIFICATE OF ANALYSIS[®]

SKU	RLS-SOL-C81
CHEMICAL NAME	CONDUCTIVITY 10000 uS/cm (microSiemens)
CHEMICAL GRADE	LABORATORY GRADE
APPEARANCE	CLEAR COLOURLESS LIQUID
LOT NUMBER	RLS-281124-4
EXPIRY DATE	28/11/2026

Warranty Certificate

<i>Test</i>	<i>Specifications</i>	<i>Results</i>
CONDUCTIVITY ($\frac{\mu S}{cm}$ Max)	10000 $\frac{\mu S}{cm}$ Max	10000 $\mu S/cm$
DENSITY (g/mL)	1.1g/ml Max	0.989 g/ml
pH @20°C	5.00 – 7.00pH @20°C	6.68pH @20°C

RLS Chemical warrants conformance of its products against the relevant specification in good faith. In-house quality control procedures include representations provided by production/manufacturing source. A representative batch retention sample of the above-mentioned material has tested according to results above. Shelf life is based on current knowledge and holds only for correct storage in the original closed and sealed containers. Suitability of the product for a particular use or application is not guaranteed. It remains the responsibility of the user to determine such.

APPROVED DATE: 28/11/2024



Quality Control

**Quality Management ISO 9001:2015
 NIST traceable calibration certified.**