

CERTIFICATE OF ANALYSIS[®]

SKU	RLS-SOL-C53
SOLUTION NAME	COBALT AA STANDARD 1000 ppm (IN 0.5N HNO₃)
CHEMICAL GRADE	LABORATORY REAGENT
CAS NUMBER	7646-79-9
APPEARANCE	COLOURLESS SOLUTION
LOT NUMBER	RLS-111224-002
EXPIRY DATE	11/12/2026

Warranty Certificate

<i>Test</i>	<i>Specifications</i>	<i>Results</i>
ASSAY (ppm)	1000 ppm MAX.	1000 ppm
DENSITY (g/mL) @ 25 °C	1.00 Max (g/mL)	0.98 g/mL
BOILING POINT (°C)	100 °C	100 °C
STABILITY TEST	COMPLIES	PASS

RLS Chemical warrants conformance of its products against the relevant specification in good faith. In-house quality control procedures include representations provided by production/manufacturing source. A representative batch retention sample of the above-mentioned material has tested according to results above. Shelf life is based on current knowledge and holds only for correct storage in the original closed and sealed containers. Suitability of the product for a particular use or application is not guaranteed. It remains the responsibility of the user to determine such.



Quality Control

**Quality Management ISO 9001:2015
NIST traceable calibration certified.**

APPROVED DATE: 12/12/2024